


MK-7: Broadband RF to mm-Wave S-Parameter Measurements for Semiconductor Transistor and IC Test


Gavin Fisher and Anthony Lord




Thanks to Joint Collaboration





Need for RF testing of ALL Semiconductor Devices



A 'non-RF' Laptop contains CPU, GPU, RAM, SSD etc
All requiring RF device modelling of transistors and components





Why is Device Modeling Critical (& Process Design Kits)




IC Fabrication

N+?




Challenges of Broadband Measurements (other than calibration)




Challenges of Broadband Measurements to 220GHz

- Broadband solutions typically require
 - Multiple probes
 - Multiple extenders
 - Multiple calibrations
 - Multiple measurements
- Then the data needs stitching together
 - Discontinuities
- Whole process is time consuming, manual and intensive




Challenges of Calibration Drift

Calibrating on Aux Chuck or measuring DUT on Right side




Measuring DUT on left side




Thermally Isolated Extenders

Inclined – Thermally Isolated Extenders




- Having the extenders inclined offers naturally improved thermal isolation
- Air jets improve cooling of platen surface
- Result is extenders stay at ambient temperature and not affected by thermal chuck
- This greatly improves drift stability regardless of chuck location


Comparison Between Extender/Probe Integration



Horizontal Extender – 50mm VDI WG - Probe




Horizontal Extender – Direct Connect - Probe




Inclined Extender – 45deg WG - Probe

Drift comparison – Horizontal Extender – 50mm WG - Probe

VDI native port saver - Max EVM




	PORT1	PORT2
Max EVM	5.8	5.9
$\Delta^{\circ}\text{C}$ at Max EVM	-0.87	+0.51
Max EVM per $\Delta^{\circ}\text{C}$	6.66	11.5




Drift comparison – Horizontal Extender – Direct Connect - Probe


Direct connect with Horizontal extender




	PORT1	PORT2
Max EVM	13.16	13.6
$\Delta^{\circ}\text{C}$ at Max EVM	-2.45	-2.45
Max EVM per $\Delta^{\circ}\text{C}$	5.37	5.55




Drift comparison – Inclined Extender – 45deg WG - Probe



	PORT1	PORT2
Max EVM	3.02	3.37
$\Delta^{\circ}\text{C}$ at Max EVM	-1.01	-1.01
Max EVM per $\Delta^{\circ}\text{C}$	2.99	3.33





Measurement comparisons – Line



- Measurements are Line 4 from the MLTRL post corrected
- No obvious degradation of return loss from any of the configurations tested (a sensitive metric)
- Transmission measurements within

Measurement comparisons – Symmetric offset short





- Comparable results between all configurations
- Some phase offsets due to positioning errors
- Evidence of improved isolation when using included
 - To be investigated why...



eVue IV
Digital Imaging System


FORMFACTOR



MicroChamber®

AttoGuard® / PureLine™


New Dual Band 220GHz Integration




- Combines coax and waveguide bands via diplexer integral to the probe
- Single sweep measurements
 - One set of probes
 - One Calibration
 - One Measurement

Features & Benefits of Dual Band

- Re-use existing tools
 - Probe station, extenders, positioners and tophat enclosure
- Manual, semi-auto or fully-auto systems
- Full thermal capability
- Dark, EMI Shielded and dry measurements
- Allows an existing N5291A to be extended to 220 GHz
- High power output to the device
- Bias is provided locally via the N5291A bias tee




Minimum settable power accurate to the probe tip




- After correction of the install Cal files (@ the Probe Tip) settable to -40 dbm
- Correction applied using supplied values for probes, 130 GHz Rf cables and power table for VDI extenders
- Alternatively - data can be obtained from 2 tier probe calibration if cal kit available

Maximum settable levelled power of -6 dbm throughout the band




- Higher power levels are possible but this is the maximum that can be levelled for the band

Dynamic Range





- 10 Hz with probe in Air and chuck down

Verification of power at the probe tip..



- Power evaluated using 1mm also PM5 VDI power meter



Thermally tested to 150 °c




- Probe tip at 150 °c but flange limited to 52 °c worst case and the rest of the extender is in ambient air with platen jet cooling

System evaluation



- Approaches were taken similar to those used for our Wafer level measurement solutions namely
 - Measurements of multiple transmission lines
 - Repetitive calibrations without moving the probes and evaluating for variation of post LRRM open and Cal to cal worst case S_{ij} variation
 - Repetitive calibrations with measurement of ISS line standards
 - Measurement drift as a function of time and temperature
 - Measurements of active devices



Measurement of transmission lines T220





Comparative measurements with I145 and T220 Broadband



Calibration / Measurement repeatability


- 100 Hz IF
- 10 MHz to 220 GHz
- 40 ps with 5 cycles on 185-400 50 um specific iss with LRRM

40 ps Line standard comparison - 5 Calibration and measurement cycles




Calibration / Measurement repeatability

40 PS Repetitive Calibration and Line measurements





- A new calibration precedes each device measurement

Calibration / Measurement repeatability - Phase





- Aside from outliers Phase for a 40 ps line is repeatable within +/- 1 degree

Calibration repeatability – Open in air variation




- 10 calibration cycles with eLRRM
- -6dbm 100 Hz 801 points
- Air vents for the air conditioning system were directed away from the system

Calibration Repeatability – Error set Sij comparison




- Compared using WinCalXE Error set comparison tool
- This was automated using a sequence
- Uses same error set as Open comparison

Drift at ambient temperature



- System was calibrated and left to drift with measurements every 3 minutes at -6 dbm 100 Hz
- S par sweep taken but fixed frequency point used that was next to the crossover point
- Air temperature measured with logger

Video of drift evaluation with Wincal remoting and Python




The screenshot displays a remote Python environment. On the left, a Python console window shows the execution of a script that performs drift processing. The script includes file paths and uses `execfile` and `exec` to run external modules. A keyboard interrupt is noted, followed by a successful execution of the drift processing script.



On the right, the RF Data Viewer displays two graphs showing the percentage delta of output versus frequency (GHz). Both graphs indicate a 0.00% delta at the specified output frequencies.

Graph Title	Output Frequency (GHz)	% Delta
%DELTA PORT1	128.70	0.00
%DELTA PORT2	134.20	0.00

Real life active device data (Teledyne device)




Active demo device measured at -30 dbm



- Thanks to Rob Sloan for designing and providing the demo device...



New Modular Positioners



- Choose manual or programmable body
- Multiple arms for each application
- Fast-swap dovetail mounting – Easy & safe
- Upgradable
- RF TopHat/IceShield compatible


Allows optimized measurements, lower cost of ownership and ease of use

Storage Pod for easy and safe swapping



Values of Inclined Waveguide Connection

- Calibration Stability
 - Frequency extender is isolated from effects of thermal chuck
 - Reduces drift due to varying heat from chuck as you measure across the wafer
- Measurements over temperature
 - The only solution for full range hot and cold measurements without condensation
 - Full TopHat for frost-free, dark and EMI-shielded
- 45 deg port saver to avoid extender damage
 - When swapping probes, reduce risk of damage to extender test port
 - Improves crosstalk between ports



Values of single sweep broadband solution

- No time taken to swap bands
- Less effort combining results
- Less potential for mismatch between bands
- Full over temperature solution 10 MHz to 220 GHz
- Full power levelling -40 dBm to -6dBm
- Power calibrated to the probe tip