


# Cascade Impedance Standard Substrate Map

000111100010

➤ **P/N: 129-246**

Pitch: 100  $\mu$ m - 150  $\mu$ m




Configuration: **GSSG, GSS, SSG, GS**



## Key to Map

### Key to the 129-246 Map



Substrate specifications: Material: Alumina; Thickness: 25 mils (635  $\mu\text{m}$ ); Dielectric constant: 9.9

 <p><b>Thru</b> Delay: 2.0ps</p>	 <p>Or</p>  <p><b>Cross Thru</b> Delay: 2.3ps</p>	 <p><b>Load</b></p>  <p><b>Precision 50 <math>\Omega</math> Load</b></p>	<p>For optimum calibration accuracy <b>only the Red</b>-marked load standards should be used.</p> <p><b>DC accuracy:</b> +/- 0.3 %</p> <p><b>Note:</b> Ensure the bias supply is turned <b>off</b> during calibration. Applying bias to the probe during calibration could cause the resistance of the load to change.</p>	 <p><b>Open</b></p>  <p><b>Short</b></p>	 <p><b>A-H Alignment Marks</b> See below for more information on alignment marks.</p>
---	---	--	--	--	--

	Verification Lines (from left to right)				
Thru Delay (ps)	27	7	14	3	40
Length ( $\mu\text{m}$ )	3500	900	1800	450	5250

	Overtravel
ACP	75 - 100 $\mu\text{m}$
Infinity	50 - 75 $\mu\text{m}$

All of the above specifications are based on the recommended overtravel (downward movement of probe after initial touchdown on the substrate) listed above. This amount of overtravel can be set before calibration on the Impedance Standard Substrate (ISS) using the alignment marks (allows precise setting of probe separation and overtravel). Figure 1 shows that initial contact with the edge of the probe tips should be made at reference plane X. The desired overtravel and thus skate (forward movement of probe tips after initial contact with substrate) is then achieved by adjusting the Z height on the positioner to move the edge of the probe tips to reference plane Y. This can also be seen from the photographic images shown in Figure 2.



**Calibration Coefficients** are dependent on the probe tip configuration, placement on a standard, and the standard configurations. This leads to unique calibration coefficients for a unique pair of probe and ISS. Therefore, the calibration coefficients are supplied with the probe not with the ISS.